



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert[®] CT/NG

Cepheid Catalogue Part No.:

Kit Lot No.:

Cartridge Lot No.:

Kit Expiration Date:

Legal Manufacturer

Cepheid AB
Röntgenvägen 5
SE-17154 Solna
Sweden

Manufacturing Facility

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Solna Sunnyvale

Functional Testing

| <i>Test Description</i> | <i>Acceptance Criteria</i> | <i>Test Result</i> |
|-------------------------|----------------------------------|--------------------|
| Positive | CT detected; NG detected | Passed |
| Negative | CT not detected; NG not detected | Passed |

02 AUG 19

Signature of Quality Assurance

Date

Name:

Title: