

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert	® CT/NG				
Cepheid Catalogue Part No.: GXCT/NG-CE-10					
Kit Lot No.: 1000	333892				
Cartridge Lot No.: 33:	523				
Kit Expiration Date: 2	2023-11-19	ž			
Legal Manufacturer Cepheid AB Röntgenvägen 5 SE-17154 Solna Sweden			Manufacturing Facility Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden	Solna ○	Sunnyvale
Functional Testing					
Test Description	Acceptance Criteria			Test Result	
Positive	CT detected; NG detected			Passed	
Negative	CT not detected; NG not detected			Passed	
Signature of Quality Assurance					
Signature of Quality Assurance Date					
Name: Wafa Qasim					
Title: QA Analyst					