

Title: QA Analyst

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Directive and the Canadian Medical Device Regulations (CMDR).				
Product Name: Xpert® HBV Viral Load				
Cepheid Catalogue Part No.: GXHBV-VL-CE-10				
Kit Lot No.: 1000	273470			
Cartridge Lot No.: 15	301			
Kit Expiration Date:	2022-11-20			
Legal Manufacturer		Manufacturing Facility	Solna	Sunnyvale
Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden		Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden		
Functional Testing according to D21898 Rev.: D				
Test Description	Ассері	ance Criteria		Test Result
LOW Positive	HBV DETECTED: 1.26-1.70 log10 IU/mL			Passed
HIGH Positive	HBV DETECTED: 5.54-5.86 log10 IU/mL			Passed
Negative	HBV NOT DETECTED			Passed
Ble 20210716				
Signature of Quality Assurance Date				
Name: William Palmén				