



## CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert® MTB/RIF Ultra

Cepheid Catalogue Part No.:

Kit Lot No.:

Cartridge Lot No.:

Kit Expiration Date:

Legal Manufacturer

Cepheid AB  
Röntgenvägen 5  
SE-171 54 Solna  
Sweden


Manufacturing Facility

Cepheid AB  
Röntgenvägen 5  
SE-171 54 Solna  
Sweden

Solna     Sunnyvale

*Functional Testing*

Test Description	Acceptance Criteria	Test Result
Wild Type Control	MTB DETECTED VERY LOW; Rif Resistance NOT DETECTED or MTB DETECTED LOW; Rif Resistance NOT DETECTED or MTB DETECTED MEDIUM; Rif Resistance NOT DETECTED or MTB DETECTED HIGH; Rif Resistance NOT DETECTED	Passed
Mutant Control	MTB DETECTED VERY LOW; Rif Resistance DETECTED or MTB DETECTED LOW; Rif Resistance DETECTED or MTB DETECTED MEDIUM; Rif Resistance DETECTED or MTB DETECTED HIGH; Rif Resistance DETECTED	Passed
Negative	MTB NOT DETECTED	Passed

   
Signature of Quality Assurance                      Date

Name:

Title: