

Certificate Of Analysis



Orphée S.A. assures that this reagent batch was manufactured under control of proper Quality Management System (ISO 13485).

Each batch undergoes a series of stringent validation and control tests

Product: Only One
Reference: HM22-002-1
Batch number: 1939

Use Before: 22-September-2021

Tested on:	September 24, 2019		
Tested by:	J. GIRAUD		
Tested with:	Mythic 22 Nb: 200518-002286		
Other Reagents:	ORPHEE HM22-001-1 Lot number: 904-26	Exp date: 2021-06-30	
	ORPHEE HM22-003-10 Lot number: 906-86	Exp date: 2021-07-31	
		L MYT1910	Exp date: 2019-12-05
	Hematology Controls Lot number:	N MYT1910	Exp date: 2019-12-05
		H MYT1910	Exp date: 2019-12-05
		Specifications	Conclusion
		*Osmolarity:	335±10 Pass
Physico-Chemical properties		*pH at 25°C:	6.55±0.10 Pass
		*Conductivity (mS.cm ⁻¹):	17.0±1.0 Pass
Bio-Chemical efficiency	Performance Tests:		Pass

Conclusion:

Actual-value meets specifications

Date: 25.09.2019

Prepared by Anna Spryszak / QC

Date: 25.09.2019

Approved by Robert Wiśniewski / QMS Representative

19, chemin du Champ-des-Filles – CH-1228 PLAN-LES-OUATES / GENEVA / SWITZERLAND

Phone : +4122 884 90 90 – Fax : +4122 884 90 99

WEB : www.orphee-medical.com - Email : contact@orphee-medical.com