


CERTIFICATE OF ANALYSIS

Product Name: Prolex™ Streptococcal Grouping Latex Kit	Lot Number: C11124
Product Code: PL.030	Expiration Date: 2015 02

Product Code	Quantity	Components
PL.031	1 x 3.0 ml	Streptococcal Group A Latex Reagent
PL.032	1 x 3.0 ml	Streptococcal Group B Latex Reagent
PL.033	1 x 3.0 ml	Streptococcal Group C Latex Reagent
PL.034	1 x 3.0 ml	Streptococcal Group D Latex Reagent
PL.035	1 x 3.0 ml	Streptococcal Group F Latex Reagent
PL.036	1 x 3.0 ml	Streptococcal Group G Latex Reagent
PL.037	1 x 3.2 ml	Streptococcal Extraction Reagent 1
PL.038	1 x 3.2 ml	Streptococcal Extraction Reagent 2
PL.039	2 x 8.0 ml	Streptococcal Extraction Reagent 3
PL.040	1 x 2.0 ml	Polyvalent Positive Control (A,B,C,D,F,G)
	300	Mixing Sticks
PL.092-48	48 cards	Test Cards

Additional Information	
Storage Conditions:	2-8°C
Precautions:	Refer to Instructions for Use

Quality Control Testing		
	Criteria	Result
	Chemical extraction of Strep group specific antigens.	PASS
	Strong agglutination of the test isolate when tested against the corresponding latex reagent.	PASS
	No agglutination reaction of the test isolate when tested against heterologous latex reagents.	PASS
	Strong agglutination of the Positive Control when tested against the latex reagents.	PASS

Release	
2013 11 20	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
20 Mural Street, Unit #4, Richmond Hill, Ontario,
L4B 1K3
support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
21 Cypress Blvd., Suite 1070, Round Rock,
Texas 78665-1034
ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
3 Bassendale Road, Bromborough, Wirral,
Merseyside CH62 3QL
uksupport@pro-lab.com