

### CERTIFICATE OF ANALYSIS

<b>Product Name:</b> Streptococcal Grouping Reagent Latex Suspension Group C	<b>Lot Number:</b> C14989E	
<b>Product Code:</b> PL.033	<b>Expiration Date:</b> 2018 11	


#### Additional Information

Storage Conditions:	2-8°C
Precautions:	Refer to Instructions for Use

#### Quality Control Testing

Appearance	Reaction	Result
Homogenous suspension with no agglutination	N/A	Satisfactory
<b>Organisms</b>		
<i>Streptococcus equisimilis</i> ATCC 12388 (Group C)	Positive	PASS
<i>Streptococcus</i> Group C - Clinical Isolate	Positive	PASS
<i>Streptococcus pyogenes</i> ATCC 19615 (Group A)	Negative	PASS
<i>Streptococcus agalactiae</i> ATCC 12386 (Group B)	Negative	PASS
<i>Enterococcus faecalis</i> ATCC 49477 (Group D)	Negative	PASS
<i>Streptococcus</i> sp. ATCC 12392 (Group F, type 2)	Negative	PASS
<i>Streptococcus</i> sp. ATCC 12394 (Group G)	Negative	PASS
<i>Streptococcus</i> Group D - Clinical Isolate	Negative	PASS
<i>Listeria monocytogenes</i> NCTC 11994	Negative	PASS
<b>Control</b>		
Polyvalent Positive Control (PL.040)	Positive	PASS

#### Release

2017 07 04	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206  
 20 Mural Street, Unit #4, Richmond Hill, Ontario,  
 L4B 1K3  
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450  
 21 Cypress Blvd., Suite 1070, Round Rock,  
 Texas 78665-1034  
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
 3 Bassendale Road, Bromborough, Wirral,  
 Merseyside CH62 3QL  
 uksupport@pro-lab.com