

CERTIFICATE OF ANALYSIS

Product Name: Streptococcal Grouping - Extraction Reagent 1	Lot Number: C16214A
Product Code: PL.037	Expiration Date: 2020-06-30

Additional Information


Storage Conditions:	2-8°C
Precautions:	Refer to Instructions for Use

Quality Control Testing

Organism	Reaction*	Result
<i>Streptococcus pyogenes</i> ATCC 19615 (Group A)	Positive	PASS
<i>Streptococcus agalactiae</i> ATCC 12386 (Group B)	Positive	PASS
<i>Streptococcus equisimilis</i> ATCC 12388 (Group C)	Positive	PASS
<i>Enterococcus faecalis</i> ATCC 49477 (Group D)	Positive	PASS
<i>Streptococcus</i> sp. ATCC 12392 (Group F)	Positive	PASS
<i>Streptococcus</i> sp. ATCC 12394 (Group G)	Positive	PASS

*When used in conjunction with Extraction Reagents 2 and 3 to extract group specific carbohydrate antigens and tested against homologous latex reagents.

Release

2018 06 29	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
 21 Cypress Blvd., Suite 1070, Round Rock,
 Texas 78665-1034
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com