


### CERTIFICATE OF ANALYSIS

<b>Product Name:</b> Streptococcal Grouping - Extraction Reagent 1	<b>Lot Number:</b> C18231L	
<b>Product Code:</b> PL.037	<b>Expiration Date:</b> 2022-01-31	

Additional Information	
Storage Conditions:	2-8°C
Precautions:	Refer to Instructions for Use

Quality Control Testing		
Organism	Reaction*	Result
<i>Streptococcus pyogenes</i> ATCC 19615 (Group A)	Positive	PASS
<i>Streptococcus agalactiae</i> ATCC 12386 (Group B)	Positive	PASS
<i>Streptococcus equisimilis</i> ATCC 12388 (Group C)	Positive	PASS
<i>Enterococcus faecalis</i> ATCC 49477 (Group D)	Positive	PASS
<i>Streptococcus</i> sp. ATCC 12392 (Group F)	Positive	PASS
<i>Streptococcus</i> sp. ATCC 12394 (Group G)	Positive	PASS
*When used in conjunction with Extraction Reagents 2 and 3 to extract group specific carbohydrate antigens and tested against homologous latex reagents.		

Release	
2020 04 06	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206  
 20 Mural Street, Unit #4, Richmond Hill, Ontario,  
 L4B 1K3  
[support@pro-lab.com](mailto:support@pro-lab.com)

U.S.A. ☎ 800 522 7740 Fax 800 332 0450  
 21 Cypress Blvd., Suite 1155, Round Rock,  
 Texas 78665-1046  
[support@pro-lab.us](mailto:support@pro-lab.us)

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
 3 Bassendale Road, Bromborough, Wirral,  
 Merseyside CH62 3QL  
[uksupport@pro-lab.com](mailto:uksupport@pro-lab.com)