




www.pro-lab.com

### CERTIFICATE OF ANALYSIS

<b>Product Name:</b> Microbank™ – Mixed		<b>Lot Number:</b> C11978M
<b>Product Code:</b> PL.170/M		<b>Expiration Date:</b> 2016 08
<b>Product Code</b>	<b>Quantity</b>	<b>Components</b>
PL.170/M	80	Cryovials containing approximately 25 beads and 1 ml of cryopreservative
	1	Retrieval device
	1	Numbered storage container
<b>Additional Information</b>		
Storage Conditions:	15-30°C	
Precautions:	Refer to Instructions for Use	
<b>Quality Control Testing</b>		
	<b>Criteria</b>	<b>Result</b>
<b>Sterility</b>	Conforms	PASS
<b>Visual Inspection</b>	Conforms	PASS
	<b>Recovery of Organism</b>	<b>Result</b>
<b>Organism</b>		
<i>Escherichia coli</i> ATCC 25922	Satisfactory	PASS
<i>Staphylococcus aureus</i> ATCC 25923	Satisfactory	PASS
<i>Klebsiella pneumoniae</i> ATCC 13883	Satisfactory	PASS
<i>Clostridium perfringens</i> ATCC 13124	Satisfactory	PASS
<i>Neisseria gonorrhoeae</i> ATCC 43069	Satisfactory	PASS
<b>Release</b>		
2014 09 22		
Date	Quality Control	

Canada ☎ 800 268 2341 Fax 905 731 0206  
 20 Mural Street, Unit #4, Richmond Hill, Ontario,  
 L4B 1K3  
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450  
 21 Cypress Blvd., Suite 1070, Round Rock,  
 Texas 78665-1034  
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
 3 Bassendale Road, Bromborough, Wirral,  
 Merseyside CH62 3QL  
 uksupport@pro-lab.com