


CERTIFICATE OF ANALYSIS

Product Name: TestOxidase™	Lot Number: C11667	
Product Code: PL.390	Expiration Date: 2015 02	

Additional Information	
Storage Conditions:	15-30°C
Precautions:	Refer to Instructions for Use

Quality Control Testing		
Appearance		Result
Clear/Cloudy Colourless Solution		Satisfactory
Organism	Reaction	Result
<i>Escherichia coli</i> ATCC 25922	Negative	PASS
<i>Pseudomonas aeruginosa</i> ATCC 27853 (NCTC 12934)	Positive	PASS
<i>Pseudomonas aeruginosa</i> ATCC 25668 (NCTC 10662)	Positive	PASS

Release	
2014 05 12	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
 21 Cypress Blvd., Suite 1070, Round Rock,
 Texas 78665-1034
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com