



www.pro-lab.com

CERTIFICATE OF ANALYSIS

Product Name: IMMERSION OIL	Lot Number: 125063
Product Code: PL.396	Expiration Date: 2026 08 15

Additional Information

Description: Viscous liquid.	Production date: 2021/08/18	Storage conditions: 15-25°C Precautions: refer to SDS
------------------------------	-----------------------------	--

Quality Control Testing

QC Results:
 Appearance: Clear viscous liquid
 Refractive index at 23°C: 1.5179
 Density at 23°C: 0.92 g/ml

Test results:
 n/a

Comments:
 We hereby certify that samples of the above Lot have been tested in accordance with current specifications and that the above is a true record of the results obtained.

Release

2021/08/18	<i>mawr</i>
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424
 21 Cypress Blvd., Suite 1155, Round Rock,
 Texas 78665-1046
 support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com