


CERTIFICATE OF ANALYSIS

Product Name: <i>E. coli</i> O157 Antiserum	Lot Number: C16035	
Product Code: PL.6250	Expiration Date: 2019 10	

Additional Information	
Storage Conditions:	2-8°C
Precautions:	Refer to Instructions for Use

Quality Control Testing		
Organisms	Reaction	Result
<i>E. coli</i> O157:H7 ATCC 35150	Positive	PASS
<i>E. coli</i> O157:H7 ATCC 43895	Positive	PASS
<i>E. coli</i> O157:H- Clinical Isolates	Positive	PASS
<i>E. coli</i> Non-O157 serotype Clinical Isolates	Negative	PASS
<i>E. hermanii</i> Clinical Isolates	Negative*	PASS
*Cross reactions may be present.		

Release	
2018 05 15	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
 21 Cypress Blvd., Suite 1070, Round Rock,
 Texas 78665-1034
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com