



www.pro-lab.com

CERTIFICATE OF ANALYSIS

<b>Product Name:</b> CRYSTAL VIOLET	<b>Lot Number:</b> K32973
<b>Product Code:</b> PL.7000	<b>Expiration Date:</b> 2023 01 31

Additional Information

Description: Deep violet solution with slight odour.	Concentrate Lot number: n/a Production date: 2022/01/14	Storage conditions: 15-25°C Precautions: refer to SDS
--	--	--


Quality Control Testing

Challenged with test organisms:  
*Escherichia coli* NCTC 12241  
Oxford *Staphylococcus aureus* NCTC 12981

Test results:  
Conforms to performance specifications when used in standard Grams stain method, giving crisp staining of Gram positive and Gram negative bacteria with a clean counterstained background.

Comments:  
We hereby certify that samples of the above Lot have been tested in accordance with current specifications and that the above is a true record of the results obtained.

Release

2022/01/19	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206  
20 Mural Street, Unit #4, Richmond Hill, Ontario,  
L4B 1K3  
support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424  
21 Cypress Blvd., Suite 1155, Round Rock,  
Texas 78665-1046  
support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
3 Bassendale Road, Bromborough, Wirral,  
Merseyside CH62 3QL  
uksupport@pro-lab.com