



www.pro-lab.com

**CERTIFICATE OF ANALYSIS**

<b>Product Name:</b> RABBIT COAGULASE PLASMA	<b>Lot Number:</b> 180766
<b>Product Code:</b> PL.850-10	<b>Expiration Date:</b> 2021 07 31

**Additional Information**

Description: Rabbit plasma containing disodium EDTA as anticoagulant.	Pack size: 10ml Production date: 2018/07/25	Storage conditions: 2-8°C Precautions: refer to SDS
---	--	--

**Quality Control Testing**

Raw Materials and Components:  
All incoming chemicals, raw materials and components are quality controlled against predetermined specifications to ensure reproducibility.

Production:  
Production is carried out against a Product Manufacturing Procedure incorporating Standard Operating Procedures and Good Manufacturing Practices.

Finished Product:  
Lot to lot comparisons are carried out to ensure that the appearance of the material, as well as the specificity and sensitivity, remains constant.

Quality Assurance  
Testing against S.aureus NCTC 12981 / ATCC 25923, S.aureus NCTC 6571 / ATCC 9144 and S.epidermidis NCTC 13360 / ATCC 12228 is carried out against predetermined parameters. The above lot passed all tests.

Stability and Shelf Life of Lyophilised Product  
Recommended storage temperature:2-8°C. Storage at lower temperatures has no detrimental effect. Extended stability trials at 4°C, 25°C and 37°C have shown that the product has a minimum shelf life of 2 years at 4°C and at least one month at 25°C. Reconstituted material, using sterile distilled water, should not be stored longer than necessary and should never exceed the recommendations of the product literature.

**Release**

2018/07/25	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206  
20 Mural Street, Unit #4, Richmond Hill, Ontario,  
L4B 1K3  
support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450  
21 Cypress Blvd., Suite 1070, Round Rock,  
Texas 78665-1034  
ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
3 Bassendale Road, Bromborough, Wirral,  
Merseyside CH62 3QL  
uksupport@pro-lab.com