

## QUALITY CONTROL CERTIFICATE

### **KPC + MBL Confirm ID Pack, Neo-Sensitabs**

REF: 98006  
LOT: 1205-1  
Date of analysis: 02-2012, 05-2012, 11-2012, 05-2012  
Date of expiry: 2015-02

Appearance: Light yellow, plane tablets, 9 mm in diameter, on both sides printed with a black identification code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:


Test strains	Results
<i>E. coli</i> ATCC 25922	Negative
<i>K. pneumoniae</i> NCTC 13438	KPC Positive
<i>K. pneumoniae</i> ATCC BAA – 1705	KPC Positive
<i>K. pneumoniae</i> ATCC BAA – 2146	MBL Positive
<i>K. pneumoniae</i> NCTC 13439	MBL Positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

ROSCO DIAGNOSTICA A/S



José B. Casals  
Quality Assurance Manager