

QUALITY CONTROL CERTIFICATE

ESBL + AmpC Screen ID pack, Neo-Sensitabs

REF: 98008

LOT: 1403-1

Date of expiry: 03-2016

Appearance: Light yellow, plane tablets, 9 mm in diameter, on both sides printed with a black identification code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:


Test strains	Results
<i>E. coli</i> ATCC 25922	Negative
<i>K. pneumoniae</i> ATCC700603	ESBL Positive
<i>Ent. cloaceae</i> NCTC 13406	AmpC positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

ROSCO DIAGNOSTICA A/S



José B. Casals
Quality Assurance Manager