

## QUALITY CONTROL CERTIFICATE

### **ESBL + AmpC Screen ID pack, Neo-Sensitabs**

REF: 98008

LOT: 2002-1

Date of expiry: 2022-01

Appearance: Light yellow, plane tablets, 9 mm in diameter, on both sides printed with a black identification code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains	Results
<i>E. coli</i> ATCC 25922	Negative
<i>K. pneumoniae</i> ATCC700603	ESBL Positive
<i>Ent. cloaceae</i> NCTC 13406	AmpC positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than  $\pm 1$  mm from mean.  
Complies.

Rosco Diagnostica A/S



Peter Samuelsen  
Manager Quality Assurance