

## QUALITY CONTROL CERTIFICATE

### KPC, MBL and OXA-48 Kit, Neo-Sensitabs

REF: 98015

LOT: 2020-1

Date of expiry: 2021-07

Appearance: Light yellow, plane tablets, 9 mm in diameter, on both sides printed with a black identification code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains	Results
<i>E. coli</i> ATCC 25922	Negative
<i>K. pneumoniae</i> NCTC 13438	KPC Positive
<i>K. pneumoniae</i> ATCC BAA – 1705	KPC Positive
<i>K. pneumoniae</i> ATCC BAA – 2146	MBL Positive
<i>K. pneumoniae</i> NCTC 13439	MBL Positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelsen  
Manager Quality Assurance