

QUALITY CONTROL CERTIFICATE

Total Metallo-beta-lactamasas Confirm Kit

REF: 98016

LOT: 2101-1

Date of expiry: 2023-07

Appearance: Light yellow, plane tablets, 9 mm in diameter, on both sides printed with a black identification code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains	Results
<i>E. coli</i> ATCC 25922	Negative
<i>K. pneumoniae</i> ATCC BAA - 2146	MBL Positive
<i>K. pneumoniae</i> NCTC 13438	MBL Positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelson
Manager Quality Assurance