



CERTIFICATE OF ANALYSIS

ESR_COA_SI305#1734A.pdf

Product Code	Name	Quantity manufactured	Type of Tubes
SI 305.100-A	LATEX CONTROLS KIT - provetta Greiner6 TESTS	100	Greiner
SI 305.300-A	LATEX CONTROLS KIT - provetta Greiner30 TESTS	156	Greiner

Lot Number	Manufacturing Date	Expiry Date		
1734A	10/10/2016	08/04/2017		

Standard Values			Measured Values				Internal Aggregation	Volume checked		
Transmittance	ESR (mm/h) (min-Val-max)		ESR (mm/h) (min-Val-max)		Instrument	Serial Number	OK			
Level 2	1534	6 - 11	9	9	9	TEST 1 BCL	2173	<input checked="" type="checkbox"/>	NO	YES
Level 3	980	15 - 22	19	19	20	TEST1 THL	PRDX	<input checked="" type="checkbox"/>	NO	YES
Level 4	637	56 - 74	65	66	68	ROLLER 20 PN	133	<input checked="" type="checkbox"/>	NO	YES

Date of Control **12/10/2016**

Date of Approval **12/10/2016**

Controlled By **Cirio Alberto**

Approved By **Pividori Denis**

Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy
Tel (+39) 049 0992000 - Fax (+39) 049 5855434
Mail PEC: info.alifax@pec.alifax.com
info@alifax.com - www.alifax.com



Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy
Tel (+39) 0432 547454 - Fax (+39) 0432 547378
Mail PEC: info.udine@pec.alifax.com
info.udine@alifax.com