



CERTIFICATE OF ANALYSIS

ESR_COA_SI305#1951A.pdf

Product Code SI 305.300-A	Name Latex Controls (30 Tests)	Quantity manufactured 178	Type of Tubes Greiner
-------------------------------------	--	-------------------------------------	---------------------------------

Lot Number 1951A	Manufacturing Date 11/12/2018	Expiry Date 09/06/2019		
----------------------------	---	----------------------------------	--	--

Standard Values			Measured Values				Internal Aggregation	Volume checked		
Transmittance	ESR (mm/h) (min-Val-max)		ESR (mm/h) (min-Val-max)		Instrument	Serial Number	OK			
Level 2	1534	6 - 11	8	9	9	TEST 1 BCL	2173	<input checked="" type="checkbox"/>	NO	YES
Level 3	980	15 - 22	18	19	20	TEST1 THL	PRDX	<input checked="" type="checkbox"/>	NO	YES
Level 4	637	56 - 74	63	66	67	ROLLER 20 PN	133	<input checked="" type="checkbox"/>	NO	YES

Controlled By	Cirio Alberto	Date of Control	12/12/2018
Approved By Manufacturing Manager	Pividori Denis	Date of Approval	12/12/2018
Released By Quality	Falcone Francesco	Date of Release	12/12/2018

Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy
 Tel (+39) 049 0992000 - Fax (+39) 049 5855434
 Mail PEC: info.alifax@pec.alifax.com
info@alifax.com - www.alifax.com



Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy
 Tel (+39) 0432 547454 - Fax (+39) 0432 547378
 Mail PEC: info.udine@pec.alifax.com
info.udine@alifax.com