



# CERTIFICATE OF ANALYSIS

MIC\_COA\_SI390900#SCR18224.pdf

Product Code <b>SI 390.900</b>	Name <b>KIT URO QUICK SCREENING 360 TEST</b>	Quantity manufactured <b>25</b>	
-----------------------------------	-------------------------------------------------	------------------------------------	--

Lot Number <b>SCR18224</b>	Manufacturing Date <b>21/11/2018</b>	Expiry Date <b>20/01/2020</b>	CAP color <b>Green</b>	Printed Band <b>White</b>
-------------------------------	-----------------------------------------	----------------------------------	---------------------------	------------------------------

## Eugonic Broth

Lot Number **SCR18224**  
 Expiry Date **20/01/2020**  
 CAP color **Green**  
 Printed Band **White**

	Physical parameters <i>(Volume, transparency, colour )</i>	Decontamination at 37°C after 72 hours	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV% <= 35%	CV% <= 35%
Found	OK	OK	17.0%	< 5%

Check of the presence of the magnetic stirrer in each vial contained in the kit: **YES**

Storage Conditions: **min 4°C - Max 30°C**

Compliance: **YES**

Controlled By .....	<b>Cirio Alberto</b>	Date of Control	<b>27/11/2018</b>
Approved By Manufacturing Manager	<b>Pividori Denis</b>	Date of Approval	<b>27/11/2018</b>
Released By Quality .....	<b>Falcone Francesco</b>	Date of Release	<b>27/11/2018</b>

### Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy  
 Tel (+39) 049 0992000 - Fax (+39) 049 5855434  
 Mail PEC: info.alifax@pec.alifax.com  
[info@alifax.com](mailto:info@alifax.com) - [www.alifax.com](http://www.alifax.com)



### Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy  
 Tel (+39) 0432 547454 - Fax (+39) 0432 547378  
 Mail PEC: info.udine@pec.alifax.com  
[info.udine@alifax.com](mailto:info.udine@alifax.com)