



CERTIFICATE OF ANALYSIS

MIC_COA_SI390900#SCR20017.pdf

Product Code SI 390.900	Name KIT URO QUICK SCREENING 360 TEST	Quantity manufactured 27	
-----------------------------------	---	------------------------------------	--

Lot Number SCR20017	Manufacturing Date 29/01/2020	Expiry Date 29/03/2021	CAP color Green	Printed Band White
-------------------------------	---	----------------------------------	---------------------------	------------------------------

Eugonic Broth

Lot Number **SCR20017**
 Expiry Date **29/03/2021**
 CAP color **Green**
 Printed Band **White**

	Physical parameters (Volume, transparency, colour)	Decontamination at 37°C after 72 hours	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV% <= 35%	CV% <= 35%
Found	OK	OK	15.0%	< 5%

Check of the presence of the magnetic stirrer in each vial contained in the kit: **YES**

Storage Conditions: **min 4°C - Max 30°C**

Compliance: **YES**

Controlled By	Cirio Alberto	Date of Control	04/02/2020
Approved By Manufacturing Manager	Pividori Denis	Date of Approval	04/02/2020
Released By Quality	Cordaro Francesca	Date of Release	05/02/2020

Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy
 Tel (+39) 049 0992000 - Fax (+39) 049 5855434
 Mail PEC: info.alifax@pec.alifax.com
info@alifax.com - www.alifax.com



Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy
 Tel (+39) 0432 547454 - Fax (+39) 0432 547378
 Mail PEC: info.udine@pec.alifax.com
info.udine@alifax.com